



Neomate Infant Restraint System



*Read this Manual
and Retain for
Future Reference*

Ferno Technical Support

Customer service and product support are important aspects of each Ferno product. Please have the product serial number available when calling, and include it in all written communications. For technical support questions:

Telephone (Toll-free)	1.800.733.3766 ext. 1010
Telephone	1.937.382.1451 ext. 1010
Email	quality.products@ferno.com

Ferno Customer Relations

For ordering assistance or general information:

CANADA AND THE U.S.A.

Telephone (Toll-free)	1.877.733.0911
Telephone	1.937.382.1451
Fax (Toll-free)	1.888.388.1349
Fax	1.937.382.1191
Internet	www.ferno.com

ALL OTHER LOCATIONS

For assistance or information, please contact your Ferno distributor. If you do not have a Ferno distributor, please contact Ferno Customer Relations:

Ferno-Washington, Inc.

70 Weil Way

Wilmington, Ohio 45177-9371, U.S.A.

Telephone	+1.937.382.1451
Fax	+1.937.382.6569
Internet	www.ferno.com

EUROPEAN REPRESENTATIVE

Ferno (UK) Limited

Stubs Beck Lane, Cleckheaton

West Yorkshire BD19 4TZ, United Kingdom

Telephone	+44 (0) 1274 851999
Fax	+44 (0) 1274 851111
Internet	www.ferno.co.uk

Disclaimer

This manual contains general instructions for the use, operation and care of this product. The instructions are not all-inclusive. Safe and proper use of this product is solely at the discretion of the user. Safety information is included as a service to the user. All other safety measures taken by the user should be within and under consideration of applicable regulations. It is recommended that training on the proper use of this product be provided before using this product in an actual situation.


Retain this manual for future reference. Include it with the product in the event of transfer to new users. Additional free copies are available upon request from Customer Relations.

Proprietary Notice

The information disclosed in this manual is the property of Ferno-Washington, Inc., Wilmington, Ohio, USA. Ferno-Washington, Inc. reserves all patent rights, proprietary design rights, manufacturing rights, reproduction use rights, and sales use rights thereto, and to any article disclosed therein except to the extent those rights are expressly granted to others or where not applicable to vendor proprietary parts.

Limited Warranty Statement

The products sold by Ferno are covered by a limited warranty, which is printed on all Ferno invoices. The complete terms and conditions of the limited warranty, and the limitations of liability and disclaimers, are also available upon request by calling Ferno at 1.800.733.3766 or 1.937.382.1451.



General Information

This product is to be used by trained medical staff only. The Neomate is designed for the emergency transport of infants weighing between 3.2 kg and 6.4 kg or 7 to 14 lbs who are deemed by the medical professional, too small to be transported by using the Pedimate. All care should be taken by the medical professional in determining if this product is correct for the safe securement of the infant during transport, and whether or not an alternative conveyance device (such as a commercially tested infant car seat) should be used instead.

The Neomate has been dynamically tested on a Ferno cot/trolley to meet CEN standards for crash retention. It should be firmly secured to the cot/trolley before the infant is secured within it. The top grey restraint should be looped around the cot/trolley backrest with the buckles correctly engaged and the slack webbing tightened by using the restraint adjuster. The two lower grey restraints should be looped around each main frame rails near the siderails. Both restraints should then be tightened by using the adjusters.

The Neomate should be positioned on the cot/trolley so that when the backrest is raised slightly, the infant is sitting in a semi-upright position and not laying supine.

Once the device is properly positioned and secured, the infant can be placed in the Neomate using the familiar five point harness system that comes over the child's shoulders and under the crotch region.

Additional horizontal restraints (with plastic buckles) are provided to "hug" the infant and provide lateral support and containment for the child. The EMT/Paramedic should determine if these horizontal restraints may cause issues (in instances of labored breathing or upper body injuries) and decide on their use.

SAFETY INFORMATION

Warning

Warning notices indicate a potentially hazardous situation which, if not avoided, could result in injury or death.

 WARNING
<p>Untrained operators can cause injury or be injured. Permit only trained personnel to operate the Neomate.</p> <p>Improper use of the Neomate can cause injury. Use the Neomate only for the purpose described in this manual.</p> <p>Improper application can cause injury. Apply the Neomate only as described in this manual.</p> <p>An unattended patient can be injured. Stay with the patient at all times.</p> <p>Secure and snug all straps. Unsecured or loose straps may allow movement and result in injury or death of the patient.</p> <p>Improper maintenance can cause injury. Maintain the Neomate only as described in this manual.</p>

Important

Important notices emphasize important usage or maintenance information.

<p>Important</p>

Bloodborne Disease Notice

To reduce the risk of exposure to bloodborne diseases such as HIV-1 and hepatitis when using the Neomate, follow the disinfecting and cleaning instructions in this manual.

Symbol Glossary

The symbols defined below are used on the Neomate and/or in this users' manual. Ferno uses symbols recognized by the International Standards Organization (ISO), American National Standards Institute (ANSI) and the emergency medical services industry.



General Warning of Potential Injury



Read the Users' Manual



Load Limit



Operation Requires One Trained Operator

OPERATOR SKILLS AND TRAINING

Skills

Operators using the Neomate need:

- a working knowledge of emergency patient-handling procedures.
- the ability to assist the patient.

Training

Operator need to:

- read and understand this manual.
- be trained on the use of the Neomate.
- practice with the Neomate before using it with a patient.
- record their training information. A sample training record sheet is provided on page 13.



WARNING

Untrained operators can cause injury or be injured. Permit only trained personnel to operate the Neomate.

ABOUT THE NEOMATE

Description

The Neomate is designed to secure infants and toddlers from 7 to 14 lbs (3.2 to 6.4 kg) on a Ferno cot.

USE

The Neomate is designed for use only in an emergency setting and only by suitably trained personnel. Where child restraint is needed outside of this setting, the transport vehicle should be fitted with restraints in accordance with applicable local standards and regulations, including but not limited to, the United States Federal Motor Vehicle Safety Standards and Regulations.

The Neomate is for professional use by one or more trained operators.

The Neomate is not designed as an immobilization device and should not be used to immobilize the patient, or as part of an immobilization system.

FEATURES

The Neomate blanket is soft, lightweight vinyl. One-piece webbing straps with quick-release buckles attach it to Ferno cot frames at three points.

The five-point patient harness consists of two continuous-loop shoulder/torso straps and a crotch strap. All secure in a central, fast-release buckle and all are fully adjustable.

COMPATIBLE COTS

The Neomate is designed for use with the following Ferno cot models:

26 Series	Mondial® RS2 Stretcher
28-Z PROFlexx®	Mondial® RS3 Stretcher
29-M	Mondial® RS4 Stretcher
30-SC	EFX 1 Stretcher Chair
35-A Series	EFX 2 Stretcher
35-A+Series	EFX 3 Stretcher
35-P PROFlexx® Series	EFX MonoBloc
35-X PROFlexx® Series	4140 Series Stretchers
93-ES Series	Falcon Range
93-EX Series	Hawk Range
93-P PROFlexx® Series	Pegasus Range
93-H PROFlexx® Series	CCT Range
POWERFlexx®	Megasus

Note: The Neomate is not compatible with Models 28 or 2033 chair cots.



WARNING

Improper use of the Neomate can cause injury. Use the Neomate only for the purpose described in this manual.

LOAD LIMIT



If the load limit has been exceeded (see *Inspecting the Neomate*, page 12).

General Specifications

Open, Flat

Width	10.5 in
Length	21 in

Folded

Length	13 in
Width	10.5 in
Height	3.5 in

Weight	2.6 lb
---------------	--------

Load Limit

Minimum	7 lb (3.2 kg)
Maximum	14 lb (6.4 kg)


Webbing Rating	3500 lb (1.6 mt)
-----------------------	------------------

Most general specifications are rounded to the nearest whole number. Metric conversions are calculated before rounding the Imperial measurements.

For more information about the Neomate contact Ferno Customer Relations (page 2) or your Ferno distributor.

Ferno reserves the right to change specifications without notice.

Components



Not Shown

Hook-and-loop storage straps on reverse surface of Neomate are not visible in this view.

USING THE NEOMATE

Before Placing the Neomate in Service

- Personnel who will work with the Neomate need to read this manual and practice using the Neomate.
- Confirm that the Neomate operates properly. Follow instructions in *Inspecting the Neomate, page 12*.

General Guidelines for Use

- Medical advice is beyond the parameters of this manual.
- It is the Neomate users' responsibility to ensure safe practices for the patient and themselves.
- It is the responsibility of the qualified Emergency Medical Service personnel to assess the patient's condition and determine the proper equipment and procedures to use.
- Follow standard emergency patient-handling procedures when using the Neomate.
- Stay with the patient at all times.
- Always use the patient harness.
- Use the Neomate only when all elements are in good working order.
- Use the Neomate only with the approved cot models listed on page 6.
- Never use the Neomate in a captain's chair or front seat of an ambulance.
- Never use the Neomate as an immobilization device or as part of an immobilization system.
- If the Neomate is in use at the time of a vehicular accident, remove the Neomate from service and replace it.

 **WARNING**

Improper application can cause injury. Apply the Neomate only as described in this manual.

 **WARNING**

An unattended patient can be injured. Stay with the patient at all times.

 **WARNING**

Secure and snug all straps. Unsecured or loose straps may allow movement and result in injury or death of the patient.

Positioning the Neomate

1. Remove any restraints attached to the cot.
2. Raise the cot backrest and lock in place at an angle between 15 and 45 degrees. This will keep the patient's shoulders higher than the pelvis and maintain the proper center of gravity.
3. Unroll the Neomate and spread it flat on the cot mattress with all straps extended.
4. Center the blanket left to right on the mattress.
5. Position the blanket with the grey backrest strap at the point where you expect the patient's shoulders to rest (Figure 1).



Figure 1 - Neomate Positioned on a Cot

6. Run the ends of the backrest strap around the cot backrest until they meet in the back, then fasten the buckle (Figure 2). Leave some slack in the strap for final adjustment.




Figure 2 - Attach the Neomate Backrest Strap

Securing the Neomate

1. Place the patient on the Neomate. If the grey backrest strap is not at the patient's shoulder level, adjust the blanket position.
2. With the blanket properly positioned, tighten the backrest strap by pulling very firmly on the free end of the strap until the mattress is compressed.
3. Fasten a main frame strap by threading the free end downward between the cot main frame and the mattress next to the head-end sidearm casting.
4. Wrap the strap up around the cot main frame and fasten the buckle (Figure 3). Leave a little slack in the strap for final adjustment.
5. Repeat with the other main-frame strap.
6. Tighten each main frame strap by holding onto the buckle with one hand and pulling firmly on the free end of the strap.




Figure 3 - Attach a Strap to the Cot Main Frame

Note: To loosen a main-frame strap, unfasten it, then grasp the buckle tang, and pull outward. Refasten the buckle.

Securing the Patient

WARNING

Secure and snug all straps. Unsecured or loose straps may allow movement and result in injury or death of the patient.

1. Pull the crotch strap buckle up between the patient's legs and lay the strap on the patient's abdomen.
2. Lift a shoulder strap over one shoulder of the patient. Place patient's arm through the strap, then lock the buckle half into the central buckle (Figure 4).
3. Repeat with the other shoulder strap.
4. Thread the shoulder strap on the patient's left side through the chest clip (Figure 5) and slide the chest clip to armpit level.




Figure 4 - Fastening a Shoulder Strap




Figure 5 - Fastening a Shoulder Strap

5. To snug the shoulder/torso straps, refer to Figure 6 and use the following procedure:
 - Snug the shoulder strap (A) against the shoulder and chest by pulling section B of the strap with one hand while steadying the central buckle with the other hand.
 - Repeat with the other shoulder strap.
 - Snug the torso strap (B) by pulling on section C with one hand while steadying the central buckle with the other hand.
 - Repeat with the other torso strap.
6. Ensure the crotch strap is fastened, crotch strap is of fixed length not adjustable




Figure 6 - Fastening a Shoulder Strap

LOOSENING THE HARNESS STRAPS

To loosen the straps, refer to Figures 7 and 8 and use the following procedures:

1. Grasp the adjustment tab (Figure 7) with one hand and pull the torso strap (B) back through the tab with your other hand.
2. Steady the central buckle with one hand and loosen the shoulder strap (A) by pulling it upward with the other hand.
3. Repeat with the other shoulder/torso strap.




Figure 7 - Loosening the Strap

Releasing the Patient from the Neomate Harness

1. Loosen the crotch strap.
2. Slide the chest clip down toward the central buckle and unthread the shoulder strap on the patient's left side from the clip.
3. Press the red release button in the central buckle and pull the buckle slightly downward to release the shoulder straps from the buckle.
4. Loosen the shoulder straps and slide them away from the patient's body.
5. Lift the patient from the Neomate according to local protocols.

Storing the Neomate

1. Detach the Neomate from the cot and spread it flat on the cot mattress.
2. Attach the shoulder straps to the central buckle and adjust all straps to lay smoothly against the blanket.
3. Fold the backrest straps across the blanket.
4. Fold the crotch panel up toward the center of the blanket.
5. Fold the cot main-frame straps and the lower blanket panels across the center of the Neomate.
6. Beginning at the crotch panel, fold the Neomate upward. Secure the roll with the hook-and-loop straps.




Figure 8 - Loosening the Crotch Strap

5 - MAINTENANCE

Maintenance Schedule

The Neomate requires regular maintenance. Set up and follow a maintenance schedule (see maintenance record sheet on page 14). The table at right represents minimum intervals for maintenance.

When using maintenance products, follow the manufacturers' directions and read the manufacturers' material safety data sheets.

Disinfecting the Neomate

Wipe or spray disinfectant on all Neomate and surfaces and straps. Follow the disinfectant manufacturer's directions for application and contact time.

Cleaning the Neomate

Hand wash the Neomate blanket and straps with warm, soapy water and a clean cloth or soft brush. Rinse with clear water. Dry the blanket with a towel and allow the straps to air dry.

Do not immerse the buckles in water.

Inspecting the Neomate

Have your services equipment maintenance personnel inspect the Neomate regularly. Follow the inspection checklist.

If the inspection shows damage or excessive wear, remove the Neomate from service. Ferno recommends that in addition to scheduled maintenance, operators perform a visual inspection each time the Neomate is used. If damage is noted, remove the Neomate from service immediately and replace it.

INSPECTION CHECKLIST

- Are all components present?
- Is the Neomate free of excessive wear?
- Is all stitching secure?
- Do the central buckle and adjustment tabs operate properly and smoothly?
- Are there cuts or tears in the blanket or straps?

 WARNING
Improper maintenance can cause injury. Maintain the Neomate only as described in this manual.

Minimum Maintenance Intervals	Each Use	As Needed	Each Month
Disinfecting (this page)	•		
Cleaning (this page)		•	
Inspecting (this page)		•	•

Important
Disinfectants and cleaners containing bleach, phenolics, or iodines can damage the Neomate. Do not use products containing these chemicals.

Important
If the Neomate is in use at the time of a vehicular accident, remove the Neomate from service and replace it.

U.S.A. NOTICE

The cot is for use with ambulances that meet the requirements of the "Star of Life" certification via Federal Ambulance Specification KKK-A-1822. For information, contact:

Federal Supply Services, Specifications Section
 Suite 8100
 470 E. L'Enfant Plaza, SW
 Washington, DC 204

